

Product information sheet

ERGOMETRINE

1. General information

This document is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31 [1] and Eurachem / CITAC Guides [2,3].

2. Description of the Reference Material (RM)

| | |
|-----------------------------------|---|
| Name: | Ergometrine |
| CAS number: | 60-79-7 |
| Catalog number: | 8519.19-100-5dried |
| Lot #: | 1000037135 |
| Certificate version: | 1 |
| Expiry date dried down standard: | 06.02.2026 |
| Starting material: | Ergometrine maleate, Lot# IV12602EH001 |
| Physical description of RM: | Thin film, dried down standard |
| Packaging and amount of RM: | Amber glass ampoules fitted with teflon faced butyl septa and PP screw caps 5 mL |
| Amount of RM: | 0.5 mg dried down, 100.8 µg/mL after reconstitution with 5 mL solvent |
| Name and address of the supplier: | Chiron AS Stiklestadvn. 1, N-7041 Trondheim, Norway Tel +47 73 87 44 90, Fax +47 73 87 44 99 www.chiron.no |

2.1 Intended use of the RM

- for laboratory use only
- calibration of analytical instruments

2.2 Reconstitution instruction

The standard that you have received may appear at first glance, as an empty vial. The target compound (s) is (are) in a film at the bottom of the vial. Do not open the vial until you are ready to reconstitute.

To reconstitute this RM use the following procedure:

1. Add 5 mL ± 0.014 mL of 90/10 acetonitrile/water with a graduated syringe or a volumetric pipette
2. Cap vial tightly
3. **Mix vigorously on a vortex mixer and repeat for several times over a longer period** or sonicate at room temperature
4. Always keep vial tightly capped.
5. Store the reconstituted standard at -20°C in a dark environment for **max. 8 weeks.**
6. **Store immediately at -20°C after usage to avoid degradation**



Note: Thorough mixing and sufficient mixing time is required to ensure complete reconstitution of the dried down standard!
Ergot alkaloids are highly sensitive to temperature, light and oxygen!

Product information sheet

2.3 Instruction for the correct use of the RM

The dried down standard should be stored at 2-8°C in a dark place. Before usage of the RM, the ampoules should be allowed to warm to room temperature. The recommended minimum sub-sample amount for all kinds of application is 100 µL. The expiry date of this RM is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/packages.

2.4 Hazardous situation

The normal laboratory safety precautions should be observed when working with this RM. Further details for the handling of this RM are available as safety data sheet (SDS).

3. Certified values and their uncertainties

| Ergometrine | | |
|---|---------------------------------|--------------------------|
| Compound | Mass concentration ^a | |
| | Certified value ^b | Uncertainty ^c |
| Ergometrine | 100.8 µg/mL | ± 1.4 µg/mL |
| ^a After reconstitution with 5 mL solvent Values are based on preparation data and confirmed experimentally by HPLC-FLD ^b Mass concentration based on weighed amount, purity and dilution step ^c Expanded uncertainty U (k = 2) of the value u _c according to GUM [4] | | |

3.1 Calculation of uncertainty

The uncertainty of the calibrant solution was calculated on the basis of preparation [5].

| Uncertainty components | Description | Standard uncertainty (u) | |
|--|---|---|---------------------------------|
| Purity (P) of solid Ergometrine maleate (the uncertainty of the purity corresponds to the standard deviation of repeated measurements) | P = 98.0 ± 1.0 % | u (P) = 0.6 % | a |
| Weighing procedure weighted sample: m _{ws} = 27.916 mg correction factor F _m = 0.737 for Ergometrine maleate | U(m) = 0.0019 mg + 9.20 * 10 ⁻⁶ * m _{Toxin} u(m) = U(m)/2 | u (m) = 0.001 mg | b |
| Dilution procedure volumetric flask: V _f = 200 mL pipette: V _p = 5 mL | calibration: 200 mL ± 0.15 mL repeatability: 0.02 mL volume expansion solvent calibration pipette: 5 mL ± 0.005 mL volume expansion solvent pipette | u (cal) = 0.06 mL u (rep) = 0.02 mL u (Vol. exp.) = 0.5 mL u (V) = 0.5 mL u (cal2) = 0.002 mL u (Vol. exp.2) = 0.012 mL u (V_p) = 0.01 mL | c d e f g h i |
| ^a Maximum tolerance of purity (rectangular distribution) was divided by $\sqrt{3}$ ^b Calculation of this u-value is based upon the uncertainty formula for the weighed amount as given in the calibration report from annual balance calibration ^{c, g} A triangular distribution (division by $\sqrt{6}$) was chosen for the calculation of u (cal) ^d Based on a series of ten fill and weigh experiments on a typical 200 mL flask; the value was used directly as a standard deviation ^{e, h} Based on the density of 0.7857 g/cm ³ at temperature T = 20°C and a maximum temperature variation of ± 3°C, of volume expansion, relative volume expansion coefficient of acetonitrile is 1370 * 10 ⁻⁶ /°C [6], volume expansion term (rectangular distribution) was divided by $\sqrt{3}$ ^{f, i} The three contributions are combined to give the u (V) = $\sqrt{u(cal)^2 + u(rep)^2 + u(Vol.exp.)^2}$ | | | |

Product information sheet

Calculation of the combined uncertainty u_c and the expanded standard uncertainty U

$$c_{Toxin} = \frac{10 \times m_{ws} \times P \times F_m}{V_f} = \frac{10 \times 27.916 \times 98.0 \times 0.737}{200} = 100.8 \text{ mg/L}$$

$$\frac{u_c(c_{Toxin})}{c_{Toxin}} = \sqrt{\left[\frac{(P)}{P}\right]^2 + \left[\frac{(m)}{m_{ws}}\right]^2 + \left[\frac{u(V)}{V_f}\right]^2 + \left[\frac{u(V_p)}{V_p}\right]^2} = \sqrt{\left[\frac{0.6}{98.0}\right]^2 + \left[\frac{0.001}{27.916}\right]^2 + \left[\frac{0.5}{200}\right]^2 + \left[\frac{0.01}{5}\right]^2} = 0.007$$

$$u(c_{Toxin}) = c_{Toxin} \times 0.007 = 100.8 \times 0.007 = 0.7 \text{ mg/L}$$

Calculation of expanded standard uncertainty U using a coverage factor $k = 2$

$$(C_{Toxin}) = u_c(c_{Toxin}) \times 2 = 0.7 \times 2 = 1.4 \text{ mg/L} = 1.4 \mu\text{g/mL}$$

4. Discussion of traceability:

This calibrant is certified on the basis of gravimetric preparation [5]. Thus the certified value (mass concentration of Ergometrine) is based on the weighed amount of the starting material and is therefore traceable to the stated purity of the solid raw material. High purity material represents a practical realization of concentration units, through conversion of mass to molar quantity.

5. Confirmation of certified value by HPLC-FLD:

The certified concentration of Ergometrine of the gravimetric prepared solution was confirmed by HPLC-FLD against an independently prepared reference batch.

| | | |
|------------------|---|-------------|
| column | Phenomenex Luna 3 μ Phenyl-Hexyl, 150 x 4.60 mm | |
| Sample dilution | undiluted | |
| injection volume | 2 μ L sample | |
| solvent A | 200mg/L ammonium carbonate in water | |
| solvent B | acetonitrile | |
| flow rate | 1.0 mL / min | |
| gradient | time in minutes (min) | % solvent B |
| | 0 – 5 | 50 |
| | 5 – 12.5 | 50 - 70 |
| | 12.5 – 14.5 | 70 |
| | 14.5 – 14.6 | 70 – 50 |
| | 14.6 - 20 | 50 |
| FLD | $\lambda_{EX} = 330 \text{ nm}$, $\lambda_{EM} = 415 \text{ nm}$ | |




Figure 1: HPLC-FLD chromatogram of Ergometrine

| | time [min] | area | height | concentration ^a |
|-------------|------------|-----------|-----------|----------------------------|
| Ergometrine | 2.143 | 6.205E+06 | 6.474E+07 | 101.3 \pm 3.0 μ g/mL |

^a Mean of 6 replicate measurements against reference batch, confidence interval with P = 95 %

6. Further information

The purchaser must determine the suitability of this product for its particular use. Chiron AS makes no warranty of any kind, express or implied, other than its products meet all quality control standards set by Chiron AS. We do not guarantee that the product can be used for a special application.

This document has been computer generated and is valid without a signature.



Product information sheet

References:

- [1] ISO Guide 31:2015 - 1-18, "Reference materials – contents of certificates, labels and accompanying documentation"
- [2] Eurachem / CITAC Guide, 1-37, (2003), "Traceability in Chemical Measurement"
- [3] Eurachem / CITAC Guide CG4, 1-133, (QUAM:2012.P1), "Quantifying Uncertainty in Analytical Measurement", 3rd Ed
- [4] International Organization for Standardization (ISO), (2008), "Guide to the expression of uncertainty in measurement", (GUM 1995 with minor corrections) 1st Ed. Geneva, Switzerland
- [5] R.D. Josephs, R. Krska, S. MacDonald, P. Wilson, H. Pettersson, Journal of AOAC International **86**, 50-60, (2003), "Preparation of a Calibrant as Certified Reference Material for Determination of the Fusarium Mycotoxin Zearalenone"
- [6] E.W. Flick, (1998), "Industrial Solvents Handbook", 5th Ed., Noyes Data Corp. Westwood NJ